

## **Corrigendum for GCMS specification (Dept. of Chemistry, NIT Dugapur)**

**BID REFERENCE: NITD/CHEMISTRY/DST-FIST/2016-17/01      Date: 12.12.2016**

1. The 'Thermal Conductivity Detector' mentioned in the 7<sup>th</sup> row should not be considered as essential item. TCD should only be considered as one of the optional accessories, as mentioned in 18<sup>th</sup> row of the table for the GCMS.

2. The (ii) point of 2<sup>nd</sup> column of 17<sup>th</sup> row of the table to be changed/modified as

"Consumable spare parts for five years (the list at least include Injection port glass liners – 10 Nos, GC Septum – 200 Nos, Column inlet/outlet nuts – 10 each, inlet/interface ferrules – 20 each, source insulator, vacuum oil (5 times the oil capacity of turbo molecular pump))"

### **in place of**

"Consumable spare parts for five years (the list at least include source filaments – 20 Nos, Source Heater – 20 Nos, Injection port glass liners – 10 Nos, GC Septum – 200 Nos, Column inlet/outlet nuts – 10 each, inlet/interface ferrules – 20 each, source insulator, vacuum oil (5 times the oil capacity of turbo molecular pump))".

3. The following spare parts will be included in the optional accessories and should be read as (viii) point in the 2<sup>nd</sup> column of 18<sup>th</sup> row of the table:

Source filament: 6 Nos.  
Packed column injector  
Packed column adaptor