

# National Institute of Technology Durgapur

## Centre of Excellence in Advanced Materials

Email: [coe.am@nitdgp.ac.in](mailto:coe.am@nitdgp.ac.in), Phone: 0343 275 9188

### Corrigendum

**Date: 12.12.2020**

to the Notice inviting Tender for a Glove Box with  
**Tender no. NITD/CoE/EQ/2020-21/2 dated 02.12.2020**

Following changes are made to the original specifications:

Existing specification	Revised specification
Dimension: 1200mm(L)X <b>800</b> mm(D)X 800mm(H)x 3mm thickness or more	Dimension: 1200mm(L)X <b>750mm(D)</b> X 800mm(H)x 3mm thickness or more.
Main Antechamber x 1 (Right Side) Antechamber size: <b>375</b> mm( $\phi$ ) x 550mm(L) or more	Main Antechamber x 1 (Right Side) Antechamber size: <b>370mm(<math>\phi</math>)</b> x 550mm(L) or more
The successful bidder must submit a valid bank guarantee on any nationalized bank amounting to <b>5%</b> of the order value towards performance security valid for the period of warranty plus 90 days. (In page 1)	The successful bidder must submit a valid bank guarantee on any nationalized bank amounting to <b>10%</b> of the order value towards performance security valid for the period of warranty plus 90 days. (In page 1)

Additional Requirements to be fulfilled by the bidder:

- The vendor should provide the contact details including phone number and email ID of at least ten end-users from Govt. agency/institute/ university to whom they have supplied the equipment. Model no. and year of supply should also be mentioned. **In addition, Performance Certificate from at least one such user should be provided.**
- If the product offered by a bidder is from an Indian OEM, the OEM should have ISO certification.
- Original Invoice, Original Warranty Certificate, Original Test Reports from OEM should be provided for all imported items/components at the time of supply of the equipment.
- Catalogue for the system and its different components/accessories like pressure gauges, sensors, controllers, etc. should be provided along with the Technical Bid.
- CE or equivalent International Certification should be provided for the unit.

All other specifications remain unchanged from original tender document as advertised on 02.12.2020.